



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 8:51 am, Nov 05, 2013

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 030451	PRINTER SN 95.1111.053	DATE OF INSPECTION 11/04/2013
LOCATION OF INSTRUMENT (STREET AND CITY) 501 FARAON ST. JOSEPH		TIME OF INSPECTION 1719

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☒ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

☒ TEMPERATURE OF ALCO SENSOR (1 O-C - 40-C)

☒ PRINTER WORKING PROPERLY

☒ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☐ SIMULATOR SOLUTION

☒ COMPRESSED ETHANOL-GAS MIXTURE

☒ STANDARD SUPPLIER AIR GAS LOT # AG300201 EXP. DATE 01/02/2015

☐ SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIMULATOR SN SIMULATOR EXP DATE

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .102

TEST 2 .101

TEST 3 .102

☒ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 1

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME
SCOTT GARY CHRIS McBANE

TYPE II PERM NUMBER EXPIRATION DATE
220179 08/03/2013 220199 08/13/2014

TELEPHONE NUMBER
816-271-5359

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901



Airgas Mid America (LABORATORY)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 2-Jan-2013

Dear Sir,

This is your Certificate of Analysis:

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
1/2/2015	108	Ethanol	0.100 ± 2% BrAC (272 ppm)
		Nitrogen	Balance

Lot # AG300201

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

<u>Serial No.</u>	<u>Concentration</u>	<u>Serial No.</u>	<u>Concentration</u>
EB0010581	391.5 ppm	EB0010603	390.9 ppm
EB0010570	258.4 ppm	EB0010559	258.3 ppm
EB0010285	208.9 ppm	EB0010595	209.2 ppm
EB0010561	101.9 ppm	EB0010562	104.9 ppm
EB0010681	53.0 ppm	EB0010579	52.4 ppm

Analytical Method: NDIR

Analyst: 

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01

AS IV Serial no: 000451
Version no: 7410

TEST RECORD 00400

Temp Date Time 9/
11/04/13 17:19 .000

Callibration Check:
20 11/04/13 17:19 .102

Subject Name

Subject I.D.

Operator Name: I.D.

Location

AS IV Serial no: 000451
Version no: 7410

TEST RECORD 00400

Temp Date Time 9/
11/04/13 17:22 .000

Callibration Check:
20 11/04/13 17:22 .101

Subject Name

Subject I.D.

Operator Name: I.D.

Location

AS IV Serial no: 000451
Version no: 7410

TEST RECORD 00404

Temp Date Time 9/
11/04/13 17:24 .000

Callibration Check:
20 11/04/13 17:24 .102

Subject Name

Subject I.D.

Operator Name: I.D.

Location

AS IV Serial no: 000451
Version no: 7410

TEST RECORD 00406

Temp Date Time 9/
11/04/13 17:26 .000

Callibration Check:
20 11/04/13 17:26 .102

Subject Name

Subject I.D.

Operator Name: I.D.

Location

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



SCOTT GARY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER; ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 08/03/2012

Number 220179

Expires 08/03/2014

MO 593-0771 (7-88)

Director of State Public Health Laboratory

Director, Department of Health

Lab. 4 (R7-88)

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



CHRISTOPHER MCBANE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER; ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 08/13/2012
Number 220199
Expires 08/13/2014

Director of State Public Health Laboratory

Director, Department of Health